



Certificate

Certificate No.: C226031.02	Date of Issue: 03 May 2021
Applicant Name:	Optoacoustics Ltd.
Applicant Address	17 Hanotea Street, Mazor, 7316000, Israel
Manufacturer (Name, Address)	Same as Applicant
Factory (Name, Address)	Same as Applicant
Product	Wireless Multi – channel Communication System uses diffused IR transmission for MRI
Model / Type Ref.	IMROC-IR
Rating	EOU 8000 - In: 100-240Vac, 50/60 Hz, 1.6A Max. Control Console - In:100-240Vac, 50/60 Hz, 0.6A Max. IR wall Detector (WD) – In: 100-240Vac, 50/60 Hz, 0.3A Max. Switch Unit and Headset - 7.4Vdc, 3000mAh (powered from rechargeable Li-ion battery)
A sample product was tested and found to be in conformity with:	IEC 60601-1-2-2014, 4 th Edition
Part of the certificate is the Test Report No.:	E226020.01

This is to certify that the product specified herein has been found to be compliant with the requirements noted above

Name: D. Shidlowsky

Signature: 

Certification Department

Notes:

