



Certificate

Certificate No.: C226030.01

Date of Issue: 2 May 2021

Applicant Name: Optoacoustics Ltd.

Applicant Address Hanotea St. 17, Mazor, Israel

Manufacturer (Name, Address) Same as applicant

Factory (Name, Address) Same as applicant

Note – place additional factories on an additional page

Product Wireless Multi – channel Communication System uses diffused IR transmission for MRI

Model / Type Ref. IMROC-IR

Rating

EOU 8000:	In: 100-240Vac, 50/60 Hz, 1.6A Max.
Control Console:	In: 100-240Vac, 50/60 Hz, 0.6A Max.
IR Wall Detector:	In: 100-240Vac, 50/60 Hz, 0.3A Max.
Switch Unit and Headset:	7.4Vdc, 3000mAh (powered from rechargeable Li-ion battery)
Battery Charger:	In: 100-240V, 50/60 Hz, 4A Max.; Out: 12Vdc, 15A, 180W Max.

A sample product was tested and found to be in conformity with:

IEC 60601-1: 2005 + CORR. 1:2006 + CORR. 2:2007 + AM1:2012; IEC 62471:2006; IEC 60601-1-6:2010, AMD1:2013, AMD2:2020; EN 50581:2012

Part of the certificate is the

Test Reports No.: S226010.01, S226011.01, S226012.01, S228840.01

This is to certify that the product specified herein has been found to be compliant with the requirements noted above

Slava Pilyagin

