



# Certificate

**Certificate No.:** C242560.01

**Date of Issue:** 8 December 2021

**Applicant Name:**

Optoacoustics Ltd.

**Applicant Address**

17 Hanotea Street, Mazor, 7316000, Israel

**Manufacturer (Name, Address)**

Same as applicant

**Factory (Name, Address)**

Same as applicant

Note – place additional factories on an additional page

**Product**

Real-time active noise-cancelling headset that enables two-way communications between the MRI Patient and MRI Staff

**Model / Type Ref.**

OptoACTIVE

**Rating**

EOU 7000 - In: 100-240Vac, 50/60 Hz, 1A Max;  
OptoACTIVE Control Console – In: 12Vdc, 2A (from AC/DC adapter);  
FOMRI-III – In: 12Vdc, 2A (from AC/DC adapter)

**A sample product was tested and found to be in conformity with:**

IEC 60601-1:2005, AMD1:2012

**Part of the certificate is the Test Report No.:**

S235860.01

---

**This is to certify that the product specified herein has been found to be compliant with the requirements noted above**

---

**Name:**

Slava Pilyagin

**Signature:**

**Certification Department**

**Notes:**

